

Body Clock Health Care Ltd

George Lane, 108
London, E18 1AD
UK

02/05/2023

Confirmation Letter Reference: CLNB1639 GBPC240660

To whom it may concern,

Confirmation of receipt of a formal application and conclusion of written agreement in the framework of Regulation EU 2023/607 amending Regulations (EU) 2017/745 and (EU) 2017/746 as regards the transitional provisions for certain medical devices and in vitro diagnostic medical devices

This letter confirms that, SGS Belgium NV, a Notified Body (NB) designated against Regulation (EU) 2017/745 (MDR) and identified by the number 1639 on NANDO, has received a formal application in accordance with Section 4.3, first subparagraph of Annex VII of MDR and has signed a written agreement in accordance with Section 4.3, second subparagraph of Annex VII of MDR with the following manufacturer:

Body Clock Health Care Ltd

George Lane, 108
London, E18 1AD
UK

Eu rep:

MedEnvoy Global B.V.

Prinses Margrietplantsoen 33, Suite
123, 2595 AM, The Hague,
The Netherlands

The devices covered by the formal application and the written agreement mentioned above are shown at the end of this letter.

In the case of devices covered by certificates issued under Directive 90/385/EEC (AIMDD) or Directive 93/42/EEC (MDD) that expired prior to the publication of the Regulation EU 2023/607 on 15 March 2023, this letter also confirms that:

- the manufacturer submitted the MDR application and signed the written agreement by the date of MDD/AIMDD certificate expiry;

The transition timelines that apply to the devices covered by this letter, subject to the manufacturer's continued compliance to the other conditions specified in Article 120.3 of MDR (as amended by EU 2023/607), are shown below:

- 26 May 2026 for Class III custom-made implantable devices
- 31 December 2027 for Class III devices and Class IIb implantable devices excluding Well-established technologies (WET - sutures, staples, dental fillings, dental braces, tooth crowns, screws, wedges, plates, wires, pins, clips and connectors)
- 31 December 2028 for other Class IIb devices, Class IIa, Class I devices placed on the market in sterile condition or have a measuring function
- 31 December 2028 for devices not requiring the involvement of a notified body under MDD but requiring it under MDR (e.g., class I devices that qualify as re-usable surgical instruments)

On behalf of the Notified Body SGS Belgium NV 1639,



Pp [Sean Kelly]
 Virginie SILORET
 Global Medical Device Certification Manager
 Email: Virginie.siloret@sgs.com
 Phone : +41 22 739 98 58

Devices covered by this letter:

Device name / Basic UDI-DI (under MDR application)	MDR Device classification (as proposed by the manufacturer and verified at the pre-application stage)	If the MDR device is a substitute device, identification of the corresponding MDD/AIMDD device	MDD/AIMDD Certificate Reference(s) of the devices under MDR application, and the NB Identification
Transcutaneous Electrical Nerve Stimulator (TENS) and Electro-Acupuncture (EA) devices for pain management and muscle re-education.	Class IIa	N/A	GB19/964595; NB1639

Confirmation Letter Revision History

Date	NB internal reference traceable to each version of the letter	Action



2023/05/02	Version 1	Initial issue
------------	-----------	---------------

SGS NB1639 - Confirmation letter Regulation (EU) 2023/607

SGS Belgium NV

Certification and Business Enhancement Registered Office: Noorderlaan 87 BE-2030 Antwerpen t +32 (0)3 545 48 48 f +32 (0)3 545 48 49
Boulevard International/Internationalelaan 55D BE-1070 Brussels t+32 (0)2 556 00 40 f +32 (0)3 545 48 49 www.be.sgs.com

Member of the SGS Group

RPR Antwerpen VAT BE 0404 882 750 Belfius 550-3560000-93